BIDOVS

THE FIRST DIALYZER WITH A HEPARIN-GRAFTED MEMBRANE

The Gambro EVODIAL dialyzer represents our specialized dialyzer for patients with a high risk of bleeding.^{1,2} It has been designed with the HeprAN heparin-grafted membrane.^{3,4} This dialyzer provides a convenient solution for patients requiring reduced and heparin-free dialysis.^{1,a,b}

FACILITATES HEPARIN-FREE DIALYSIS

- Increases the rate of successful heparin-free HD therapy sessions, compared to the current standard of care in high bleeding risk patients^{1,a}
- Allows for a reduction of systemic heparin dosing, without compromising the dialysis session⁵⁻⁷
- No heparin is released from the membrane during the dialysis session⁸

TREAT ACROSS A VARIETY OF MODALITIES

- EVODIAL dialyzers can be used in hemodialysis, hemodiafiltration and hemofiltration therapies to potentially provide individualized patient treatments⁹
- Also available for acetate-free biofiltration (AFB & AFB-K)

HIGH PERFORMING MEMBRANE®

- Anticoagulation benefits for high bleeding risk patients⁴
- Improved adsorptive capabilities vs. other types of membranes³





TYPICAL PATIENT PROFILE: **HIGH BLEEDING RISK PATIENTS AND PATIENTS WITH HEPARIN SENSITIVITIES (EXCLUDING HIT SYNDROME TYPE II)**

^a Standard of care includes saline flushes and blood predilution

^b Excluding patients with HIT (heparin-induced thrombocytopenia)

The AN 69 membrane is the predecessor to the HEPRAN membrane

Gambro EVODIAL Dialyzer HeprAN Membrane

IN VITRO CLEARANCES* (ml/min)	EVODIAL 1.0	EVODIAL 1.3	EVODIAL 1.6	EVODIAL 2.2
$Q_B = 200 \text{ ml/min}, Q_D = 500 \text{ ml/min}, Q_{UF} = 0 \text{ ml/min}$				
Urea	173	181	189	195
Creatinine	156	166	176	184
Phosphate	135	146	156	168
Vitamin B ₁₂	85	96	111	126
$Q_B = 300 \text{ ml/min}, Q_D = 500 \text{ ml/min}, Q_{UF} = 0 \text{ ml/min}$				
Jrea	216	231	250	265
Creatinine	187	204	220	237
Phosphate	156	172	187	207
Vitamin B ₁₂	92	106	124	143
$Q_B = 400 \text{ ml/min}, Q_D = 500 \text{ ml/min}, Q_{UF} = 0 \text{ ml/min}$				
Urea	241	261	287	310
Creatinine	205	226	246	269
Phosphate	168	187	205	230
Vitamin B ₁₂	96	111	131	153
ULTRAFILTRATION COEFFICIENT** (± 20 %) – ml/(h x mmHg)	33	40	50	65
Membrane				
Surface area (m²)	1.05	1.30	1.65	2.15
nner thickness (µm)			42	
nternal diameter (µm)	210			
Sieving coefficients***		_		
nulin	0.96			
Albumin	< 0.01			
Vaterials				
Membrane	HeprAN: AN 69**** membrane heparin grafted			
Housing	Polycarbonate			
Blood headers	Polycarbonate (no joint)			
Potting	Polyurethane			
Storilization		Came	mo rov	
Sterilization	G.C.		ma ray	100
Blood priming volume (ml) ±10%	66	83	100	129
Residual blood volume (ml)	< 1 450			
Maximum TMP (mmHg)	150, 200			200 500
Recommended blood flow rates (ml/min)	150-300	200-400	200-500	300-500
Weight (g)	160	200	240	255
Quantity per box	24			

^{*} $\pm 10\%$ excepted for vit.B $_{12}$ (\pm 20%) – Phosphates values measured with a pH = 7.4 \pm 0.1.

Evodial is contraindicated in patients presenting with a known allergy to heparin or who have type II thrombocytopenia caused by heparin (HIT Syndrome type II).

For the safe and proper use of the Evodial Dialyzers refer to the contraindications, warnings, precautions and the complete directions for use.



^{**} Bovine blood (Htc: 32%, protein concentration = 60 g/l, T = 37°C, QB = 300 ml/min).

^{***} Protein concentration = 60 g/l,QB = 300 ml/min, QUF = 20 % x QB = 60 ml/min.
*/**/*** Values measured according to ISO 8637.

^{****} Acrylonitrile and sodium methallyl sulfonate copolymer – Polyethyleneimine – Heparin grafted

Laville M, et al. Kidney Int 2014; 86:1260–1267.
 Kessler M, et al. Semin Dial 2015; 28:474–489.
 Thomas M, et al. Contrib Nephrol 2011; 173:119–129.
 Chanard J, et al. Nephrol Dial Transplant 2008; 23:2003–2009.
 Kessler M, et al. Hemodial int 2013; 17:282–293.
 Morena M, et al. Hemodialysis International 2010; 14:403–410.
 Frascá GM, et al. Ther Apher Dial 2015; 19:154–161.
 Baxter. Data on file. EVODIAL Risk Analysis Summary. 2013.
 Baxter. EVODIAL Instructions for use. 2010.